

## Certificate of Analysis (CoA)

<b>Product Name</b>	Medical Instrument Disinfectant; Glutaraldehyde %2 with Activator
<b>Trademark</b>	GBL® SANITAROSA™ GA ACT
<b>Product Ref No</b>	6002
<b>Order No</b>	436002025000
<b>Batch Number</b>	C5578
<b>Production Time</b>	01.06.2023
<b>Expiration Date</b>	01.06.2025
<b>Analysis Date</b>	01.06.2023-02.06.2023

Physical Properties		
Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Color	Colorless	+
Labeling	EN ISO 15223-1	+

Chemical Properties		
Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
pH @ 25 °C	4,00±0,40	4.12
Assay of Glutaraldehyde	2,00%±0,30%	2.10

Microbiological Properties	
Microorganism Name	Biological Activity
Pseudomonas aeruginosa ATCC 15442	+
Staphylococcus aureus ATCC 6538	+
Enterococcus hirae ATCC 10541	+
Candida albicans ATCC 10231	+
Aspergillus brasiliensis ATCC 16404	+
Mycobacterium terrae ATCC 15755	+
Mycobacterium avium ATCC 15769	+
Bacillus subtilis ATCC 6633	+

Berna GÜVENÇ ŞENTÜRK  
Biologist  
Quality Control Responsible

Öncü GÜNDÜ  
Chemist  
QMS Representative