

Certificate of Analysis (CoA)

Brand Name	Medical Device Surface Cleaner Disinfectant; Alcohol-free
Trademark	GBL® SURFACTO™ QAS
REF No	6007
Batch No	C1989
Order No	436007051000
Production Time	14.03.2023
Expiration Date	14.03.2025
Analysis Date	14.03.2023-15.03.2023

Physical Properties

Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Odor	Odorless	+
Color	Colorless	+
Labelling	EN ISO 15223-1	+

Chemical Properties

Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
pH @ 25 °C (ca)	7,2	7,200

Microbiological Properties

Microorganism Name	Biological Activity
<i>Staphylococcus aureus</i> ATCC 6538 1 min.	+
<i>Pseudomonas aeruginosa</i> ATCC 15442 1 min.	+
<i>Enterococcus hirae</i> ATCC 10541 1 min.	+
<i>Candida Albicans</i> ATCC 10231 5 min.	+
<i>Aspergillus brasiliensis</i> ATCC 16404 5 min.	+

Berna GÜVENÇ ŞENTÜRK
Biologist
Quality Control Responsible

Öncü GÜNDÜ
Chemist
QMS Representative