

Certificate of Analysis (CoA)

Product Name	Medical Instrument Disinfectant; Glutaraldehyde %2
Trademark	GBL® SANITAROSA™ GA
Product Ref No	6071
Order No	436071025000
Batch Number	C4232
Production Time	05.12.2022
Expiration Date	05.12.2024
Analysis Date	05.12.2022-06.12.2022

Physical Properties		
Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Color	Colorless	+
Labeling	EN ISO 15223-1	+

Chemical Properties		
Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
pH @ 25 °C	6,00±0,40	6,130
Density (g/cm ³)	1,00±0,01	0,9996
Assay of Glutaraldehyde	2,00%±0,30%	2.17

Microbiological Properties	
Microorganism Name	Biological Activity
Pseudomonas aeruginosa ATCC 15442	+
Staphylococcus aureus ATCC 6538	+
Enterococcus hirae ATCC 10541	+
Candida albicans ATCC 10231	+
Aspergillus brasiliensis ATCC 16404	+
Mycobacterium terrae ATCC 15755	+
Mycobacterium avium ATCC 15769	+
Bacillus subtilis ATCC 6633	+

Berna GÜVENÇ ŞENTÜRK
Biologist
Quality Control Responsible

Öncü GÜNDÜ
Chemist
QMS Representative