

## Certificate of Analysis (CoA)

<b>Product Name</b>	Surgical Instrument Disinfectant: Peracetic Acid, Concentrated
<b>Trademark</b>	GBL® PEROXY™ HLD
<b>Product Ref No</b>	6128
<b>Order No</b>	436128045000
<b>Batch Number</b>	C5580
<b>Production Time</b>	01.06.2023
<b>Expiration Date</b>	01.12.2024
<b>Analysis Date</b>	01.06.2023-02.06.2023

Physical Properties		
Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Color	Colourless	+
Labelling	EN ISO 15223-1	+

Chemical Properties		
Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
Peracetic Acid (% w/w)	5,0±0,5	5,20
Hydrogen Peroxide (% v/w)	20,0±1,0	20,80

Microbiological Properties	
Microorganism Name	Biological Activity
Candida Albicans ATCC 10231 5 min.	+
Aspergillus brasillensis ATCC 16404 5 min.	+
Bacillus spizizeni ATCC 6633 60 min.	+
Staphylococcus aureus ATCC 6538 5 min.	+
Pseudomonas aeruginosa ATCC 15442 5 min.	+
Enterococcus hirae ATCC 10541 5 min.	+
Mycobacterium terrae ATCC 15755 5 min.	+
Adenovirus type 5, strain Adenoid 75, ATCC VR-5 60 min.	+
Murine nanovirus (MNV) strain S99, RVB-651 60 min.	+
Poliovirus type I, LSc-2ab 60 min.	+

Berna GÜVENÇ ŞENTÜRK  
Biologist  
Quality Control Responsible

Öncü GÜNDÜ  
Chemist  
QMS Representative