

Certificate of Analysis (CoA)

Product Name	Enzymatic Cleaner for Surgical Instrument; Concentrated
Trademark	WARECLEAN™ EC
REF No	6160
Batch Number	C1990
Order No	436160131000
Production Time	14.03.2022
Expiration Date	14.03.2024
Analysis Date	14.03.2022-17.03.2022

Physical Properties		
Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Labelling	EN ISO 15223-1	+

Chemical Properties		
Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
pH @ 25 °C (ca)	7,2±0,3	7,255

Microbiological Properties	
Microorganism Name	Biological Activity
<i>S. aureus</i> ATCC 6538 15 min.	+
<i>P. aeruginosa</i> NCTC 6749 15 min.	+
<i>B. subtilis</i> var niger 15 min.	+

Berna GÜVENÇ ŞENTÜRK
Biologist
Quality Control Responsible

Öncü GÜNDÜ
Chemist
QMS Representative