

Certificate of Analysis (CoA)

Trade Name	Hemodialysis Machine Disinfectant; with Active Chlorinated
Brand Name	CITROTAL 100™
REF No	6338
Batch Number	C6821
Order No	436338045000
Production Time	18.10.2023
Expiration Date	18.10.2024
Analysis Date	18.10.2023

Physical Properties		
Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Labelling	EN ISO 15223-1	+

Chemical Properties		
Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
Active Chlorine (% w/v)	5,0±0,5	5,18

Microbiological Properties	
Microorganism Name	Biological Activity
<i>Staphylococcus aureus</i> ATCC 6538	+
<i>Pseudomonas aeruginosa</i> ATCC 15442	+
<i>Enterococcus hirae</i> ATCC 10541	+
<i>Candida Albicans</i> ATCC 10231	+
<i>Aspergillus brasiliensis</i> ATCC 16404	+
<i>Mycobacterium terrae</i> ATCC 15755	+