

Certificate of Analysis (CoA)

Trade Name	Concentrated Disinfectant for Treatment of Aspiration Systems
Trademark	LIQUISEPT™ ASPIRATION SA
REF No	6018
Order No	436818025000
Batch Number	B9289
Production Time	14.04.2021
Expiration Date	14.04.2023
Analysis Date	15.04.2021- 16.04.2021

Physical Properties

Quality Control Values	Theoretical Results	Results
Form	Liquid	+
Odor	Lemon	+
Labeling	EN ISO 15223-1	+

Chemical Properties

Quality Control Values	Theoretical Results	Results
Solubility	Soluble	+
pH @ 25°C (ca)	7,5±0,3	7,503

Microbiological Properties

Microorganism Name	Biological Activity
<i>Staphylococcus aureus</i> ATCC 6538 15 min.	+
<i>Pseudomonas aeruginosa</i> ATCC 15442 15 min.	+
<i>Enterococcus hirae</i> ATCC 10541 15 min.	+
<i>Candida Albicans</i> ATCC 10231 15 min.	+
<i>Aspergillus brasiliensis</i> ATCC 16404 60 min	+

Berna GÜVENÇ ŞENTÜRK
Biologist
Quality Control Responsible

Öncü GÜNDÜ
Chemist
QMS Representative